

## Adult ECG Electrodes

CLINICAL CRITERIA		
Packaging	Opening / Preparing	Simulated Clinical Use
Is the Expiry Date & Lot Number or Manufacturing Date & Shelf Life Clearly visible	Does the pouch / packet have a space to record date it was opened or new expiry date	Do the electrodes remove easily from the backing piece
Does the pouch clearly identify product wear time	Can the pouch / packet be easily opened without splitting or damaging the packaging using hands only	Does the product apply to the skin correctly and efficiently without wastage?
Does the pouch clearly identify the shelf life once pouch opened?	Can the pouch be resealed and is it easy to use?	Does the backing card have a colour or pattern which would make it visible on a floor or in the bedsheets.
Are there clear instructions for use, including skin-preparation procedures and electrode preparation		Is the electrode comfortable on the skin for the wearer
Is the manufacturing product code clearly displayed		Any indications / aids for removal on the product
Does the Description clearly identify the electrode properties inside?		Discomfort or skin reaction on removal
Does the pouch / packet indicate clinical use?		
	Disposal	
	Is the pouch recyclable	